

CENTRAL MEDICAL SERVICES SOCIETY
Ministry of Health & Family Welfare
(Autonomous Body under MoHFW, Govt. of India)
2nd Floor, Vishwa Yuvak Kendra, Pandit Uma Shankar Dikshit Road,
Chanakypuri, New Delhi-110021 PHONE -:011-21410905/6, Fax -:011-21410849

Date: - 17/01/2024

Minutes of Pre-bid Meeting
For Procurement of Condoms (Free Supply for NACO) for NACO
GeM Bid Reference No: GEM/2023/B/4405877, dated 28/12/2023
Pre-bid Meeting held on 03/01/2024 at 11:00 AM

1. Following officials were present during the Pre-bid meeting: -
 - (i) Dr. Shobini Rajan, DDG (NACO)- On VC
 - (ii) Mr. D Mohapatra, GM (Finance), CMSS
 - (iii) Mr. Vijay Kumar Dhingra, In-charge GM (Procurement), CMSS
 - (iv) Mr. Arvind Kumar, Associate Consultant, NACO
 - (v) Dr. Shantanu Kumar Purohit, National Consultant, NACO
 - (vi) Mr. Lava Mishra, AGM (Procurement), CMSS
 - (vii) Ms. Akanksha Jain, AGM (QA), CMSS
2. Any prospective bidders were not present during the Pre-bid meeting.
3. Queries from following prospective bidders were received via email: -

Sr. No.	Name of the bidders
1	M/s Anondita Healthcare

4. Points raised by representatives of prospective bidders were discussed. After due consideration of the queries, the Clarification/Amendments are enclosed.

ANNEXURE-1

Pre-bid queries raised by the prospective bidders & Clarification/Amendments by CMSS

Sr. No.	Tender Clause No & Pg No.	Bidder's Representation	Bidder's Name	Clarification/Amendments
1.	Annexure-1A & Pg No. 41	<p>Technical Specifications of Latex Rubber Condoms</p> <p>(The standard shall have technical specification in the line of IS/ISO 2340:2011 Reaffirmed Year 2020) to comply the provision of Rule 7 of Medical Device Rule 2017.</p> <p>IS/ISO 2340:2011 Standard is not for natural rubber latex condom</p> <p>We complies Standard ISO 4074, WHO and Schedule R for Natural Rubber Latex Condom.</p>	M/s Anondita Healthcare	<p>As per Rule 7 in Schedule Medical Devices 2017, the Product standards for medical device include:</p> <ol style="list-style-type: none">1. The medical device shall confirm to the standards laid down by the Bureau of Indian Standards established under section 3 of the Bureau of Indian Standard Act, 1985 (63 of 1985) or as may be notified by the Ministry of Health and Family Welfare in the Central Government, from time to time.2. Where no relevant Standard of any medical device has been laid down under sub-rule (1), such device shall conform to the standard laid down by the International Organisation for Standardisation (ISO) or the International Electro Technical Commission (IEC), or by any other pharmacopeial standards.3. In case of the standards which have not been specified under sub-rule (1) and sub-rule (2), the device shall conform to the validated manufacturer's standards. <p>The technical specifications of Latex Rubber Condoms require compliance to Rule 7 of Medical Device Rules 2017 that may be ascertained during technical evaluation.</p>

Note: - Above changes will be part of the tender document.

**Sd/
GM (Procurement)**

ANNEXURE-2

REVISED CRITICAL DATE SHEET ARE AS UNDER:

Description	Scheduled date
Bid Submission End Date and Time	25/01/2024 till 03:00 PM
Last date of submission of original documents	25/01/2024 till 03:00 PM
Bid Opening Date and Time	25/01/2024 at 03:30 PM

Note: - Apart from above, all other terms & conditions of tender document will remain unchanged.

**Sd/
GM (Procurement)**