

CENTRAL MEDICAL SERVICES SOCIETY

(An Autonomous Society Under Ministry of Health & Family Welfare, Govt. of India)
2nd Floor, Vishwa Yuvak Kendra, Pt. Uma Shankar Marg, Teen Murti Marg,
Chankayapuri, New Delhi-110021

Phone: 011-21410905, 21410906

Website: www.cmss.gov.in,

Amendment No. 1

Dated: 10-08-2026

Tender No.: CMSS/PROC/2026-27/NACO/027 dated: 21-07-2026

Tender Title: Global Tender for procurement of Laboratory Equipments for National AIDS Control Organization (NACO).

A. Following are the amendments made in the subject tender document: -

Schedule-1

Item 1: Microplate / ELISA Plate Reader

Sl. No.	Tender Clause / Specification	To be amended as
1.	Sr. No. 12 – Capacity of both Single and Dual wavelength reading	Capacity of Single and dual wavelength range 340 – 700 nm
2.	Facility for screen touch and LED display	Facility for screen touch and LED/LCD display
3.	Alphanumeric keypad, LED display and capacity to create, edit, or run at least 10 protocols.	Alphanumeric keypad/Menu keys for navigation, LED/LCD display and capacity to create, edit, or run at least 10 protocols
4.	Read Time-Maximum 8 Sec for 96 well plates (Single wavelength)	Read Time-Maximum 10 Sec for 96 well plates (Single wavelength)

Item 2: Microplate / ELISA Plate Washer

5.	Low wash bottle liquid volume sensor and interchangeable 8 X 12 manifolds	Standard 8 Pin Manifold and liquid volume sensor available in waste bottle.
6.	Sr. No. 5 – Fluid Delivery method: Internal positive displacement syringe / vacuum pump	Internal positive displacement syringe / vacuum pump / Diaphragm pump.
7.	S. No. 6 - Alphanumeric keypad, LED display and capacity to create, edit, or run at least 10 protocols	Alphanumeric keypad/Menu keys for navigation, LED/LCD display and capacity to create, edit, or run at least 10 protocols.



8.	Include capacity for limed shaking and soak time, programmable up to 50 min, Wash Cycle 0-6	Soaking time programmable between 0 - 999 sec, and up to 6 wash cycle.
9.	Page No. 54, ITB 9.4 – EMD amount for Schedule- 1	EMD for Schedule-I - 10,47,200.00

Schedule-2

Fully Automated and Integrated System for Nucleic Acid Extraction and Real Time (RT) PCR

10.	Point No. 22, Page No. 73 – The system should have on-board memory to store run files of at least 1500 run files	The system should have onboard memory to store run files of at least 1500 run files OR the system should feature adequate on-board data storage memory i.e. min. 50 GB to store run files.
11.	Page No. 54, ITB 9.4 – EMD amount for Schedule- 2	EMD for Schedule-II - 44,46,336.00
12.	Point No. 3, Page No. 67 – PGIMER, State-Chhattisgarh	Point No. 3, Page No. 67 – PGIMER, State-Chandigarh
13.	Point No. 2.d, Page No. 57 – The manufacturer of "Laboratory Equipments" must have supplied at least 40% of quoted quantity of same or similar item in last 02 years 2023-24 and 2024-25, including the current financial year till date of tender opening. Note: Similar item means any type of similar lab Equipment's/Machines	The manufacturer of "Laboratory Equipment's" must have supplied at least 40% of quoted quantity of same or similar item in last 02 years 2024-25 and 2025-26, including the current financial year till date of tender opening. Note: Similar item means any type of similar lab Equipment's/Machines

B. Apart from above, all other terms & conditions of the tender document remain unchanged.



Manish
10/8/26
**Joint Director
(Procurement)**

मनीष कुमार भारस्कर/Manish Kumar Bhaskar
संयुक्त निदेशक (प्रापण) / Joint Director (Procurement)
केन्द्रीय चिकित्सा सेवा सोसाइटी / Central Medical Services Society
स्वास्थ्य और परिवार कल्याण मंत्रालय / Ministry of Health and Family Welfare
भारत सरकार / Government of India